

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Digital Panel Meters
Name and address of the applicant	Entes Elektronik A.S. Dudullu OSB, 1. Cd. No 23 Umraniye ISTANBUL Turkey
Name and address of the manufacturer	Entes Elektronik A.S. Dudullu OSB, 1. Cd. No 23 Umraniye ISTANBUL Turkey
Name and address of the factory	<input type="checkbox"/> Additional information on page 2
Note: When more than one factory, please report on page 2	Entes Elektronik A.S. Dudullu OSB, 1. Cd. No 23 Umraniye ISTANBUL Turkey
Ratings and principal characteristics	MPR-1 Series, MPR-2 Series: Supply: 95 – 270 Vac/dc, 50 – 60 Hz, 5 VA or 12 - 50 Vdc, 5 VA (for D-series). Measurement Input: 10 – 300 Vac (L – N); 10 – 480 Vac (L – L), 50 – 60 Hz. MPR-3 Series: Supply: 95 – 270 Vac/dc, 50 – 60 Hz, 5 VA Measurement Input: 10 – 300 Vac (L – N); 10 – 480 Vac (L – L), 50 – 60 Hz. MPR-4 Series: Supply: 50 – 270 Vac/dc, 50 – 60 Hz, 5 VA or 24 – 60 Vac/dc, 5 VA (for D-series). Measurement Input: 5 – 300 Vac (L – N); 5 – 480 Vac (L – L), 50 – 60 Hz.
Trademark (if any)	Entes
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	Model MPR-1 Series, MPR-2 Series, MPR-3 Series, MPR-4 Series Model number may be followed by -xxx-yyyy, where x and y are alphanumeric and numeric values denoting non-critical options. (See General Product Information for explanations)

This CB Test Certificate is issued by the National Certification Body

Meander 1051, NL-6825 MJ Arnhem, Netherlands
DEKRA Certification B.V.




IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

As shown in the Test Report Ref. No. which forms part of this Certificate

Additional information on page 2

IEC 61010-1:2010, IEC 61010-2-030:2010

National differences:

CA, US

2199714.50 (consisting of subreport with suffixes A, B and C)

This CB Test Certificate is issued by the National Certification Body

Meander 1051, NL-6825 MJ Arnhem, Netherlands
DEKRA Certification B.V.



A handwritten signature in blue ink, appearing to read 'Wim Singewald', written over a white background.